

AD A109271

DTIC FILE COPY

ACOUSTIC MICROSCOPY AT CRYOGENIC TEMPERATURES

Annual Summary Report

1 July 1980 - 30 June 1981

Contract No. N00014-77-C-0412

G.L. Report No. 3369

November 1981

Reproduction in whole or in part is permitted  
for any purpose of the United States Government.

**DISTRIBUTION STATEMENT A**

Approved for public release;  
Distribution Unlimited

Edward L. Ginzton Laboratory  
W.W. Hansen Laboratories of Physics  
Stanford University  
Stanford, CA 94305

DTIC  
ELECTE  
S JAN 4 1982 D  
D

82 01 04 038

UNCLASSIFIED

SECURITY CLASSIFICATION OF THIS PAGE (When Data Entered)

REPORT DOCUMENTATION PAGE		READ INSTRUCTIONS BEFORE COMPLETING FORM
1. REPORT NUMBER	2. GOVT ACCESSION NO.	3. RECIPIENT'S CATALOG NUMBER
	AD-A109 274	
4. TITLE (and Subtitle)		5. TYPE OF REPORT & PERIOD COVERED
ACOUSTIC MICROSCOPY AT CRYOGENIC TEMPERATURES		Annual Summary Report 1 July 80 - 30 June 81
		6. PERFORMING ORG. REPORT NUMBER
		G.L. Report No. 3369
7. AUTHOR(s)		8. CONTRACT OR GRANT NUMBER(s)
C.F. Quate J. Heiserman		N00014-77-C-0412
9. PERFORMING ORGANIZATION NAME AND ADDRESS		10. PROGRAM ELEMENT, PROJECT, TASK AREA & WORK UNIT NUMBERS
Edward L. Ginzton Laboratory W.W. Hansen Laboratories of Physics Stanford University, Stanford, CA 94305		Task NR384-924
11. CONTROLLING OFFICE NAME AND ADDRESS		12. REPORT DATE
Office of Naval Research Physics Division Code 412 Arlington, Virginia 22217		November 1981
		13. NUMBER OF PAGES
		10
14. MONITORING AGENCY NAME & ADDRESS (if different from Controlling Office)		15. SECURITY CLASS. (of this report)
		UNCLASSIFIED
		15a. DECLASSIFICATION/DOWNGRADING SCHEDULE
16. DISTRIBUTION STATEMENT (of this Report)		
"Approved for public release; distribution unlimited"		
17. DISTRIBUTION STATEMENT (of the abstract entered in Block 20, if different from Report)		
NOT REPRODUCIBLE EXCEPT BY AUTHORITY OF THE NATIONAL ARCHIVES AND RECORDS ADMINISTRATION		
18. SUPPLEMENTARY NOTES		
19. KEY WORDS (Continue on reverse side if necessary and identify by block number)		
Acoustic microscopy Liquid argon Mechanical scanning Superfluid helium Liquid nitrogen High frequency acoustic properties		
20. ABSTRACT (Continue on reverse side if necessary and identify by block number)		
<p>This report summarizes our progress in a research program devoted to high resolution acoustic microscopy and cryogenic temperatures. During the past year we have studied the acoustic properties of two cryogenic fluids - liquid argon and liquid <math>^4\text{He}</math> below the temperature of <math>0.5^\circ\text{K}</math>. In argon we have determined that nonlinear effects provide us with a dramatic increase in resolving power for focused acoustic beams. In liquid helium we have measured the attenuation versus temperature of sound at 1 GHz where the wavelength is less than <math>2500 \text{ \AA}</math>. The predicted behavior below <math>0.5^\circ\text{K}</math> where the attenuation varies as the fourth</p>		

DD FORM 1 JAN 73 1473

EDITION OF 1 NOV 65 IS OBSOLETE  
S/N 0102-LF 014-6601

UNCLASSIFIED

SECURITY CLASSIFICATION OF THIS PAGE (When Data Entered)

409640

UNCLASSIFIED

SECURITY CLASSIFICATION OF THIS PAGE(When Data Entered)

power of temperature is confirmed for the first time at these small wavelengths. It gives us further confidence in our program to realize an instrument that will permit us to examine our specimens with an acoustic probe that is less than 1000 Å in diameter.

Prescription For	
NTIS GRA&I	<input checked="checked" type="checkbox"/>
DTIC TAB	<input type="checkbox"/>
Unannounced	<input type="checkbox"/>
Justification	
By	
Distribution/	
Availability Codes	
Dist	Avail and/or Special
F	

UNCLASSIFIED

SECURITY CLASSIFICATION OF THIS PAGE(When Data Entered)

## SUMMARY OF ANNUAL REPORT

Our principle goal in this research program is to explore the acoustic properties of liquid helium with wavelengths below 1000 Å. We hold the ultimate goal of using a scanned acoustic microscope to explore microscopic features of natural materials such as metals, ceramics, polymers and fabricated structures as used in microelectronic devices. The present techniques available for studying materials and structures in the submicron world are limited to the electron microscope. It is insensitive to the elastic properties of the samples. Contrary to this, the acoustic microscope operating at room temperature has proven to be a suitable method for studying the elastic properties.

We hope to gain new insight into material properties -- grains and grain boundaries in metals -- inhomogeneities in ceramics or polymers -- and defects in microelectronic devices. It will be one of the major applications of helium -- a cryogenic fluid with unusual properties.

The year covered by this report produced several advances that bring us close to the realization of an operating instrument which will allow us to explore the microscopic region that we are so interested in. We have, also, advanced our knowledge of the properties of liquid helium at temperature below 1°K as explored with sound waves near 1 GHz where the wavelength is less than 2500 Å .

One major event during the year was the Rank Foundation Conference in London in September 1980. The entire conference was focused on the principle of mechanical scanned microscopy. Major areas of microscopy were treated. It now turns out microscopy with optical, x-rays and acoustics use this principle

to generate images with valuable information. Our work was a major component of the conference. The results are now completely published by Academic Press.<sup>1</sup> We attach the title page and table of contents as an Appendix.

We have a dilution refrigerator capable of cooling the microscope to the neighborhood of 0.15°K that will now dissipate 0.2 milliwatts of power. This will take care of the power dissipated by our mechanical scanning circuits.

The attractive feature of liquid helium at this temperature is very low attenuation for acoustic waves. It is also attractive because the wave velocity is  $0.24 \times 10^{-5}$  cm/sec -- a factor of eight below water. We have in this program succeeded in putting together a form of acoustic interferometer which has permitted us to measure the acoustic attenuation at the low temperature less than 1°K. That work confirms the fact that the attenuation decreases as the fourth power of the temperature. At that temperature (where we propose to operate) the attenuation is reduced to the point where it is no longer a limitation on the operation frequency. This is in contrast to room temperature instruments where the basic limitation is the intrinsic attenuation of the liquids.

Little experimental work has been done in liquid helium since the low impedance of the liquid presents a major obstacle. The impedance of helium is 0.03 -- about 50 times lower than that of water. If we are to exploit the marvelous properties that are found there we must deal with the problem of coupling acoustic energy from a high impedance crystal to the low impedance liquid. We have used the approach of matching layers since it is well known that a layer of material one quarter wavelength in thickness will serve to eliminate reflections provided that the impedance of the layer can be made equal to the geometric mean of the high crystal impedance and the low liquid

<sup>1</sup>Scanned Image Microscopy, E. Ash, Editor (Academic Press, N.Y., 1980).

impedance. This is no easy task since the optimum impedance of the layer, near 1.0 mech-ohm, is lower than any that is readily available in solids. It has required a major effort to find the solution but we think it has now been done. The work was adequately described in our 1 st annual report (G.L. Report No. 8149, July 1980).

During the interval covered by this report we have made the decision to use single carbon layers rather than the tungsten-carbon layer. A double layer would permit more efficient transmission of energy through the interface but we would be limited to a rather narrow band of frequencies. Many carbon layers have now been fabricated on various surfaces. We think we understand the process rather well. Of major interest to this program is our newly acquired ability to deposit the carbon on the curved surfaces of the acoustic lens. We have mounted the coated lenses in helium and tested their focusing properties by translating a flat reflector through the focal plane. The reflected signal rises to a maximum and then decreases as the reflector is moved continuously through the focal region. This confirms our belief that the principles developed at room temperature for acoustic microscopy can now be translated over to the cryogenic instrument.

The component that remains to be examined is the system for mechanical scanning. We have preliminary indications that the system used in liquid argon (and nitrogen) can be adapted to helium. It will be part of the program during the first quarter of the next annual period.

One distinguished result with cryogenic fluids during this interval has been the discovery by Rugar -- Hunt Fellow for 1981 -- that the resolution can be improved by pushing the amplitude of the sound waves into the region of nonlinear behavior. The evidence for improved resolution has been demonstrated in the images as described previously.

As we now understand the phenomena it comes about as a result of the generation of second harmonic near the focal region of our spherically converging waves. These second harmonic waves are generated on spherical wavefronts and they converge to a waist (at the focus) that is narrower than that achieved by the fundamental wave. The theory has not been fully developed but it will be a component of the work for the next annual period.

Much of cryogenic microscopy have been developed through extensive work in liquid argon and nitrogen. It is easier to work there and there we have enjoyed considerable success. The images in nitrogen and argon are superior to optical images. We feel that they are merely a prologue to the work that is ahead of us in liquid helium below 1°K.

## PUBLICATIONS

D. Rugar, J. Heiserman, S. Minden and C.F. Quate, "Acoustic Micrographs of Human Metaphase Chromosomes," *Journal of Microscopy* 120, No 2, 193-199, November 1980.

C.F. Quate, "The Acoustic Microscopy: A Concept of Microscopy Using Waves of Sound," *Naval Research Reviews*, Office of Naval Research, Vol. 33, No. 1, 24-32, Fall-Winter, 1980-1981.

J. Heiserman, "Cryogenic Acoustic Microscopy," *Proceedings of the Rank Prize International Symposium on Scanned Image Microscopy*, p. 71, E. Ash, Editor, Academic Press 1980.



## APPENDIX

Proceedings of the Rank Prize International Symposium.

London, England

September 1980.

# SCANNED IMAGE MICROSCOPY

*Edited by*

ERIC A. ASH

*Department of Electrical Engineering  
University College, London*

1980



**ACADEMIC PRESS**

*A Subsidiary of Harcourt Brace Jovanovich, Publishers*  
**London New York Toronto Sydney San Francisco**

## CONTENTS

Contributors	v
Preface	ix
Fundamentals of Scanning Systems <i>G.S. Kino</i>	1
SCANNING ACOUSTIC MICROSCOPY	
Microwaves, Acoustics and Scanning Microscopy <i>C.F. Quate</i>	23
Gas Medium Acoustic Microscopy <i>H.K. Wickramasinghe and C.R. Petts</i>	57
Cryogenic Acoustic Microscopy <i>J. Heiserman</i>	71
Acoustic Microscopy: Imaging Microelectronic Circuits with Liquid Metals <i>J. Attal</i>	97
Mechanically Scanned Acoustic Microscope composed of Plane and Concave Transducers for Transmission Mode <i>N. Chubachi</i>	119
Metrology and Imaging in the Acoustic Microscope <i>R.D. Weglein</i>	127
Spectroscopic Study of Defects in Thick Specimen Using Transmission Scanning Acoustic Microscopy <i>J.K. Wang, C.S. Tsai and C.C. Lee</i>	137
Characterisation of Acoustic Microscopes for Non-destructive Testing <i>B. Nongaillard, J.M. Rouvaen, R. Torquet and E. Bridoux</i>	149

Defect Detection for Microelectronics by Acoustic Microscopy <i>R.L. Hollis and R.V. Hammer</i>	155
<b>SCANNING OPTICAL MICROSCOPY</b>	
Theory and Principles of Optical Scanning Microscopy <i>W.T. Welford</i>	165
Developments in High Resolution Confocal Scanning Light Microscopy <i>G.J. Brakenhoff, J.S. Binnerts and C.L. Woldringh</i>	183
Imaging Modes of Scanning Optical Microscopy <i>C.J.R. Sheppard</i>	201
Scanning Optical Microscopy of Semiconductor Devices <i>T. Wilson, J.N. Gannaway and C.J.R. Sheppard</i>	227
Resolution of Near-Field Optical Fibre Refractive Index Profiling Methods <i>W.J. Stewart</i>	233
A CCD Linear Array Based Scanning System for Rapid Analysis of Biomedical Material on Standard Microscope Slides <i>A.S.J. Farrow</i>	241
<b>SCANNING PHOTOACOUSTIC MICROSCOPY</b>	
Scanning Photo-Acoustic Microscopy <i>Y.H. Wong</i>	247
Photoacoustic Microscopy at Low Modulation Frequencies <i>M.V. Luukkala</i>	273
Thermal-Wave Imaging and Microscopy <i>A. Rosenwaig</i>	291
Electron-Acoustic Microscopy <i>G.S. Cargill</i>	319
Photothermal Radiometry for Spatial Mapping of Spectral and Material Properties <i>P.E. Nordal and S.O. Kanstad</i>	331

Contents	xv
The Optacoustic and Photothermal Microscope: The Instrument and its Applications <i>G. Busse</i>	341
The Mirage Effect in Photothermal Imaging <i>D. Fournier and A.C. Bocarra</i>	347
Photoacoustic Imaging of Compositional Variations in $\text{Hg}_{1-x}\text{Cd}_x\text{Te}$ Semiconductors <i>J.F. McClelland, R.N. Kniseley and J.L. Schmit</i>	353
SCANNING SOFT X-RAY MICROSCOPY	
The Scanning X-Ray Microscope - Potential Realisations and Applications <i>E. Spiller</i>	365
Imaging and Scanning Soft X-Ray Microscopy with Zone Plates <i>G. Schmahl, D. Rudolph and B. Niemann</i>	393
A Scanning Ultrasoft X-Ray Microscope with Multilayer Coated Reflection Optics: First Test with Synchrotron Radiation Around 60 eV Photon Energy <i>R.-P. Haelbich</i>	413
Photoelectron Detection of X-Ray Images for Contact Microscopy and Microanalysis <i>F. Polack</i>	435
X-Ray and Particle Microscopy using Fresnel Zone Plates <i>N.M. Ceglie</i>	443
Scanning Soft X-Ray Microscopy - First Tests with Synchrotron Radiation <i>H. Rarback, J. Kenney, J. Kirz and X.-S. Xie</i>	449
Subject Index	457